



CIRCE Building
Campus Río Ebro - University de Zaragoza
Mariano Esquillor Gómez, 15 - 50018 Zaragoza
Tel.: 976 761 863 Fax: 976 732 078
www.fcirce.es


Assessment report of power quality analyzer Circutor MYeBOX according to IEC 61000-4-30 requirements for Class A equipment

Jorge Bruna Romero / Juan José Pérez Aragüés
Area RYS
Project PC-17/0348
2018, 5th of June



ISO 9001:2008
ISO 14001:2004
OHSAS 18001:2007
ISO 50001:2011

www.tuv.com
ID 9105076991

	Document:	Assessment report of power quality analyzer Circutor MYeBOX 1500 according to IEC 61000-4-30 requirements for Class A equipment	Version:	0
	Project:	PC-17/0348	Author:	JBR / JPA
			Date:	5/6/18

1 Aim

This document presents the assessment results of power quality analyzer Circutor MYeBOX 1500 according to requirements of the edition currently in force of standard IEC 61000-4-30 for Class A equipment. This assessment is performed to guarantee the fulfilment of this analyzer accuracy requirements for the following power network conditions:

- Rated voltage: 230 V (U_{din})
- Rated frequency: 50 Hz / 60 Hz

To verify compliance with the aforementioned analyzer, calibration results obtained in the equipment serial number 083217190026 installed with firmware version **001.002.000**, by two laboratories accredited by ENAC, have been taken into account:

- Certificate number 18792, issued by CIRCUTOR
- Certificate number 4367, issued by LME-CIRCE

The following section includes the statement made by the analyzer manufacturer regarding the firmware version control installed on the equipment.

1.1 Firmware version control

Hereby, CIRCUTOR S.A, specifies that the version control of the portable power network analyzer MYeBOX is carried out as follows:

The firmware version installed in analyzer MYeBOX consists of 3 groups of 3 characters:

XXX.YYY.ZZZ

Where:

XXX - Hardware model and associated peripherals

YYY - Version of measurement algorithms


ZZZ - Version corresponding to the firmware relating to user interface: internal and external communications, display control and data delivery to APP, cloud and files

This firmware version can be visualized both on the MYeBOX display and the APP.

In the header of the files just the last two groups of characters is shown

YYY.ZZZ.

Viladecavalls, 9th of April of 2018

	Document:	Assessment report of power quality analyzer Circutor MYeBOX 1500 according to IEC 61000-4-30 requirements for Class A equipment	Version:	0
	Project:	PC-17/0348	Author:	JBR / JPA
			Date:	5/6/18

1.2 Manufacturer's certificate regarding model MYeBOX 150

This section includes the certificate issued by the manufacturer of the analyzer regarding the fact that model MYeBOX 1500 has the same structure, measurement method, hardware and firmware as model MYeBOX 150.



Web: www.circutor.com
 E-mail: central@circutor.com
 Tel: (+34) 93 745 29 00
 Fax: (+34) 93 745 29 14

**CERTIFICADO
 CERTIFICATE**

Mediante este documento CIRCUTOR, S.A. garantiza que el modelo de analizador portátil MYeBOX 1500 comparte la misma estructura, método de medida, hardware y firmware que el modelo MYeBOX 150. CIRCUTOR certifica que las únicas diferencias entre estos modelos son la entrada del canal de corriente de fugas, la entrada del canal de la tensión de referencia, las entradas/salidas transistor y las comunicaciones 3G. No hay ninguna otra diferencia que afecte a la medida y cálculo de variables eléctricas.

Through this document CIRCUTOR, S.A. guarantees that the MYeBOX 1500 portable analyzer model has the same structure, measurement method, hardware and firmware as MYeBOX 150 model. CIRCUTOR certifies that the only differences between these models are the input of the leakage current channel, the channel input of the reference voltage, the transistor inputs / outputs and the 3G communications. There is no other difference that affects the measurement and calculation of electrical variables.

Carlos Córcoles
 Responsable de Producto - División Gestión Energética
 Product Manager - Energy Management Division



CIRCUTOR, SA



2 Results

The following tables show the degree of compliance with the results of calibration performed by CIRCUTOR and LME-CIRCE, whose results are contained in certificates number 18792 and 4367, meets the requirements of the edition currently in force of standard IEC 61000-4-30, according the maximum error values or limits allowed for Class A equipment.

In all cases, compliance statement is based on a 95 % probability of coverage for the expanded uncertainty of the measurement results on which the compliance decision is based.

Magnitude	Range	Limit	Compliance
Frequency	42.5 Hz – 57.5 Hz 51 Hz – 69 Hz	10 mHz	Yes
AC voltage (50 Hz / 60 Hz)	11 V – 345 V	$0.1 \% \cdot U_{din}$	Yes
Flicker (230 V at 50 Hz / 60 Hz)	Pst 0.2 – 10	5 % or 0.05 (the highest)	Yes
Voltage dips, interruptions and swells (230 V at 50 Hz / 60 Hz)	5 % – 110 % 0.1 s – 10 s	$0.2 \% \cdot U_{din}$ 20 ms / 16 ms	Yes ¹
Voltage harmonics (230 V at 50 Hz / 60 Hz)	Order 2 nd – 50 th 100 %, 10 %, 200 %	5 % if $\geq 1 \%$ else $0.05 \% \cdot U_{din}$	Yes
Voltage unbalance Negative sequence coeff. (u_2)	0.161 % – 6.927 %	0.15 %	Yes

¹ In this case only the obtained deviation from reference is considered. It is not possible to declare compliance using a probability of 95 % coverage for the expanded uncertainty even though the result of the measurement is below the limit.

2.1 Conclusions

In view of the above results, it can be concluded that the analyzer object of the calibration meets the requirements of the standard IEC 61000-4-30:2015 for Class A under the technical conditions specified in Section 2 “Results”.

Signed: Jorge Bruna Romero Electrical Test Responsible	Signed: Juan José Pérez Aragüés Technical Expert